

Certificate Number : ECT 2026/0216- 01

ULR Number : CC2519 26 0 00002722F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Deep Freezer

Make : Remi

Range : -45 to -86 °C

Model : ULT-185

Resolution : 0.1 °C

S.No : ZEGD-24432

Accuracy : As Per Manual

ID No. : 16

Location : Microbiology Lab

Environmental Conditions

Temperature 25.5 °C

Relative Humidity 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	
1	Digital Thermometer With RTD Sensor	22000243 &T0256/1719	SAN/25/1042/TH/1	11.08.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	-69.7	-69.50	-69.50	-69.51	-69.50	-0.20	--	0.17

** End of Certificate **