



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 01

Date of Issue : 9-Jun-2026

ULR Number : CC2519 26 0 00005767F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2026/0493

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 8-Jun-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 8-Jun-2026

Andipalayam Post, Tiruchengode ,

Suggested calibration Due Date : 7-Jun-2027

Namakkal- 637 214.

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Unit Under Calibration

Physical Condition : Good

Name : Refrigerator / Freezer

Make : LG

Range : 2 to 8 °C / -18 to -20 °C

Model : Intellocool

Resolution : NA

S.No : --

Accuracy : As per Manual

ID No. : RF-SS-01

Location : Chemical Lab

Environmental Conditions

Temperature 25.5 °C

Relative Humidity 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018822 / 1711	ECT 2025/0971-01	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ±°C
		1	2	3		
1	Refrigerator 2 to 8	3.7	3.7	3.6	3.67	0.24
2	Freezer -18 to -20	-19.6	-19.7	-19.8	-19.70	0.25

** End of Certificate **



Checked By
H. Priya. K. H.

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT