



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0976- 06

ULR Number : CC2519 25 0 00010631F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

Site: S.F.No: 571,D.No.3/183A,

Periyasolipalayam, Paramathi Velur.TK

Namakkal (DT)- 637 204.

Date of Issue : 15-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0976

Receipt Date : 13-Sep-2025

Date of Calibration : 13-Sep-2025

Suggested calibration Due Date : 12-Sep-2026

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-13

Equipment Under Calibration

Physical Condition : Good

Name : IR Thermometer

Make : Mecor

Model : IRT1050P

S.No : --

ID No. : IRT-01

Range : -50 to 1050 °C

Resolution : 0.1 °C

Accuracy : As Per Manual

Location : --

Environmental Conditions

Temperature 24.8 °C

Relative Humidity 56.1 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Infrared Thermometer	160211812	RPL-2025-0727-C-007007	18.07.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	Standard Reading °C	UUC Reading in °C			UUC Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	50.0	49.9	50.0	49.9	49.93	-0.07	--	3.50
2	100.2	100.0	100.0	100.0	100.00	-0.20	--	3.50
3	200.4	200.1	200.1	200.2	200.13	-0.27	--	3.50
4	400.8	400.5	400.4	400.4	400.43	-0.37	--	3.50
5	500.8	500.3	500.3	500.4	500.33	-0.47	--	3.50

** End of Certificate **



Checked By
(Signature)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT