

EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 29

ULR Number : CC2519 25 0 00010533F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 12-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0966

Receipt Date : 10-Sep-2025

Date of Calibration : 10-Sep-2025

Suggested calibration Due Date : 9-Sep-2026

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-09

Equipment Under Calibration

Physical Condition : Good

Name : Cooling Centrifuge (RPM)

Make : Remi

Model : C-24 Plus

S.No : VCEK-6987

ID No. : IR-CC-99-01

Range : Max 20000 RPM

Resolution : 100 RPM

Accuracy : As Per Manual

Location : Chemical Lab

Environmental Conditions

Temperature 25.8 °C

Relative Humidity 57.5 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Tachometer	S622195	ECT 2024/1122-16	06.10.2025

Calibration Results

Parameter - Mechanical /Acceleration and Speed

S.No	UUC Reading RPM	Standard Reading in RPM			Standard Reading (Mean) RPM	Error RPM	Error Allowed ±RPM	Uncertainty ±RPM
		1	2	3				
1	3000	3003	3003	3003	3003.00	-3.00	--	57.86
2	5000	5005	5005	5004	5004.67	-4.67	--	57.86
3	7000	7006	7006	7007	7006.33	-6.33	--	57.86
4	10000	10009	10009	10010	10009.33	-9.33	--	57.86
5	12000	12011	12012	12013	12012.00	-12.00	--	57.87

** End of Certificate **

Checked By
(Hampiya K.H.)




Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IIInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 28

ULR Number : CC2519 25 0 00010532F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 12-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0966

Receipt Date : 10-Sep-2025

Date of Calibration : 10-Sep-2025

Suggested calibration Due Date : 9-Sep-2026

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Cooling Centrifuge (Temperature)

Make : Remi

Model : C-24 Plus

S.No : VCEK-6987

ID No. : IR-CC-99-01

Range : 2 to 40 °C

Resolution : 1 °C

Accuracy : As Per Manual

Location : Instrumentation -V

Environmental Conditions

Temperature : 25.7 °C

Relative Humidity : 57 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Calibrator with Sensor	091109171/1703	ECT 2025/0142-03	04.02.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	4	4	3.9	3.8	3.9	3.87	0.13	--	0.68
2	8	8	7.8	7.8	7.7	7.77	0.23	--	0.68
3	10	10	9.6	9.5	9.5	9.53	0.47	--	0.68
4	15	15	14.3	14.4	14.3	14.33	0.67	--	0.68

** End of Certificate **


Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT