



ECT/FT-17

EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 23

Date of Issue : 12-Sep-2025

ULR Number : CC2519 25 0 00010527F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2025/0966

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 10-Sep-2025

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 10-Sep-2025

Andipalayam Post, Tiruchengode ,

Suggested calibration Due Date : 9-Sep-2026

Namakkal- 637 214.

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Hot Air Oven-1

Make : Servo Scientific

Model : --

S.No : 1340

ID No. : IR-DO-16-01

Range : Amb to 200 °C

Resolution : 0.1 °C / 1 °C

Accuracy : ± 2 °C

Location : Chemical Lab-Instrumentation Room-I

Environmental Conditions

Temperature 25.8 °C

Relative Humidity 57 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1708	ECT 2024/1122-08	06.10.2025
2	Digital RTD Calibrator	11110002959	CT25-7-1	09.01.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	30.0	30.0	30.1	30.1	30.1	30.10	-0.10	2.00	0.70
2	35.0	35.0	35.2	35.1	35.1	35.13	-0.13	2.00	0.70
3	37.0	37.0	37.2	37.2	37.3	37.23	-0.23	2.00	0.70
4	42.0	42.0	42.4	42.4	42.5	42.43	-0.43	2.00	0.70
5	44.0	44.0	44.6	44.6	44.6	44.60	-0.60	2.00	0.70
6	45.0	45.0	45.6	45.6	45.7	45.63	-0.63	2.00	0.70
7	45.0	45.0	45.6	45.6	45.7	45.63	-0.63	2.00	0.70

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

ECT

Checked By