



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 24

ULR Number : CC2519 25 0 00010528F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 12-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0966

Receipt Date : 10-Sep-2025

Date of Calibration : 10-Sep-2025

Suggested calibration Due Date : 9-Sep-2026

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Hot Air Oven-2

Make : Servo Scientific

Model : --

S.No : 1340

ID No. : IR-DO-17-02

Range : Amb to 200 °C

Resolution : 0.1 °C/ 1 °C

Accuracy : ± 2 °C

Location : Chemical Lab-Instrumentation Room-I

Environmental Conditions

Temperature : 25.8 °C

Relative Humidity : 59 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1708	ECT 2024/1122-08	06.10.2025
2	Digital RTD Calibrator	11110002959	CT25-7-1	09.01.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	60.0	60.0	59.9	59.9	59.8	59.87	0.13	2.00	0.70
2	100	100	99.7	99.7	99.8	99.73	0.27	2.00	0.70
3	102	102	101.7	101.6	101.5	101.60	0.40	2.00	0.70
4	105	105	104.3	104.3	104.4	104.33	0.67	2.00	0.70
5	130	130	129.3	129.3	129.2	129.27	0.73	2.00	0.70
6	140	140	139.1	139.2	139.2	139.17	0.83	2.00	0.70
7	200	200	199.0	199.0	199.0	199.00	1.00	2.00	0.70

** End of Certificate **



Checked By
Hampiya.K.H.

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT