



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 21

Date of Issue : 4-Mar-2026

ULR Number : CC2519 26 0 00002742F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2026/0216

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 2-Mar-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 2-Mar-2026

Andipalayam Post, Tiruchengode,

Suggested calibration Due Date : 1-Mar-2027

Namakkal - 637 214.

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Hot Air Oven

Make : Labtech / Selec

Range : Amb to 250 °C

Model : TC303

Resolution : 1 °C

S.No : 120165

Accuracy : As Per Manual

ID No. : HO-02

Location : Chemical Lab

Environmental Conditions

Temperature 25.8 °C

Relative Humidity 59 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	60	60	59.9	59.9	59.8	59.87	0.13	--	0.68
2	80	80	79.7	79.7	79.8	79.73	0.27	--	0.68
3	102	102	101.5	101.5	101.5	101.50	0.50	--	0.68
4	105	105	104.4	104.3	104.3	104.33	0.67	--	0.68
5	130	130	129.3	129.3	129.2	129.27	0.73	--	0.68
6	140	140	139.2	139.1	139.0	139.10	0.90	--	0.68

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT