



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 26

ULR Number : CC2519 25 0 00010530F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 12-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0966

Receipt Date : 10-Sep-2025

Date of Calibration : 10-Sep-2025

Suggested calibration Due Date : 9-Sep-2026

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Cooling Water Bath

Make : Labtech

Model : PITC721

S.No : --

ID No. : IR-CW-62-01

Range : 20 to 60 °C

Resolution : 0.1 °C

Accuracy : ± 2 °C

Location : Chemical Lab-Instrumentation Room-IV

Environmental Conditions

Temperature : 25.8 °C

Relative Humidity : 59 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1707	ECT 2024/1122-07	06.10.2025
2	Digital RTD Calibrator	11110002959	CT25-7-1	09.01.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	20.0	20.2	20.2	20.2	20.20	-0.20	2.00	0.70
2	30	30.3	30.3	30.3	30.30	-0.30	2.00	0.70
3	40	40.4	40.4	40.4	40.40	-0.40	2.00	0.70
4	50	50.5	50.5	50.5	50.50	-0.50	2.00	0.70
5	60	60.5	60.6	60.6	60.57	-0.57	2.00	0.70

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT

Checked By
Hampriya.KH