



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 18
ULR Number : CC2519 26 0 00002739F
Customer Name and Address
M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 4-Mar-2026
Pages : 1 of 1
SRF Number : 2026/0216
Receipt Date : 2-Mar-2026
Date of Calibration : 2-Mar-2026
Suggested calibration Due Date : 1-Mar-2027
Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Muffle Furnace
Make : Servo Scientific
Model : --
S.No : 20251101
ID No. : IR-MF-20-02

Range : Amb to 1200 °C
Resolution : 1 °C
Accuracy : As Per Manual
Location : Chemical Lab (Instrumentation - I)

Environmental Conditions

Temperature : 26.5 °C Relative Humidity : 59 %RH Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator with Thermocouple sensor	220615018808 / TCN1510	ECT 2025/0971-06	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	300	300	300.4	300.5	300.5	300.47	-0.47	--	1.93
2	550	550	550.7	550.8	550.9	550.80	-0.80	--	1.93
3	600	600	601.3	601.3	601.4	601.33	-1.33	--	1.93
4	800	800	802.5	802.6	802.7	802.60	-2.60	--	1.93
5	1000	1000	1003.2	1003.3	1003.4	1003.30	-3.30	--	1.93

** End of Certificate **

Checked By
(Hanipriya.K.H)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT