



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 27

Date of Issue : 12-Sep-2025

ULR Number : CC2519 25 0 00010531F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2025/0966

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 10-Sep-2025

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 10-Sep-2025

Andipalayam Post, Tiruchengode ,

Suggested calibration Due Date : 9-Sep-2026

Namakkal- 637 214.

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : COD Digester

Make : Spectra Lab

Range : Amb to 200 °C

Model : COD 2015 M

Resolution : 1 °C

S.No : SB033

Accuracy : ± 2 °C

ID No. : IR-CD-32-01

Location : Chemical Lab-Instrumentation Room-I

Environmental Conditions

Temperature 25.9 °C

Relative Humidity 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1708	ECT 2024/1122-08	06.10.2025
2	Digital RTD Calibrator	11110002959	CT25-7-1	09.01.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed $\pm$ °C	Uncertainty $\pm$ °C
		1	2	3				
1	50	49.8	49.8	49.8	49.80	0.20	2.00	0.90
2	70	69.7	69.7	69.6	69.67	0.33	2.00	0.90
3	100	99.5	99.5	99.5	99.50	0.50	2.00	0.90
4	120	119.4	119.3	119.3	119.33	0.67	2.00	0.90
5	150	149.2	149.2	149.2	149.20	0.80	2.00	0.90
6	200	199.0	198.9	198.9	198.93	1.07	2.00	0.90

** End of Certificate **



Checked By
Hampriya K.H.

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT