



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0970- 02

ULR Number : CC2519 25 0 00010558F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode ,
Namakkal- 637 214.

Date of Issue : 13-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0970

Receipt Date : 11-Sep-2025

Date of Calibration : 11-Sep-2025

Suggested calibration Due Date : 10-Sep-2026

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-13

Equipment Under Calibration

Physical Condition : Good

Name : Infrared Thermometer

Make : Mecro

Model : IRT1050P

S.No : --

ID No. : CL-IT-45-01

Range : -50 to 1050 °C

Resolution : 0.1 °C

Accuracy : As Per Manual

Location : Chemical Lab

Environmental Conditions

Temperature 25.5 °C

Relative Humidity 58.5 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Infrared Thermometer	160211812	RPL-2025-0727-C-007007	18.07.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	Standard Reading °C	UUC Reading in °C			UUC Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	50.5	50.3	50.4	50.4	50.37	-0.13	--	3.50
2	100.8	100.6	100.6	100.6	100.60	-0.20	--	3.50
3	200.6	200.2	200.1	200.1	200.13	-0.47	--	3.50
4	500.3	499.1	499.2	499.2	499.17	-1.13	--	3.00

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT