



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, € Certificate Number : ECT/ NN 2026/0110 -01
Ambattur, Chennai - 600 053. Tamil Nadu. € Receipt Number : NN 2026/0110 / 2-Mar-2026
Tel : 044 - 4852 1288. € Calibration Date : 2-Mar-2026
Mobile : 9486195184, 9176032184. € Date of Issue : 4-Mar-2026 Page 1 of 1

CERTIFICATE OF CALIBRATION

Name and Address of Customer

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode
Namakkal - 637 214.

Customer Reference

Work Order

Calibrated at : Onsite

Equipment Under Calibration

Name : Conductivity Meter
Make : Eutech
Model : CON700
S.No : 2723785
ID No : WL-CM-12-01

Range : 0 to 200 μ S/cm
Resolution : 0.01/0.1/1 μ S/cm
0.01/0.1 ms
Accuracy : As Per Manual
Location : Chemical Lab

Enviromental Condition :

Temperture : 25.7 $^{\circ}$ C
Humidity : 58 % RH

Calibration Procedure : ECT/WI-NN-01

Physical condition Good

Standard Equipment used for Calibration

S.No	Description	Product Code	Batch No	Master Validity
1	84 us/cm@ 25 $^{\circ}$ C	ECCON84BT	335/01	15.08.2026
2	100 us/cm@ 25 $^{\circ}$ C	ECCON100BT	442/02	06.06.2026
3	500 us/cm@ 25 $^{\circ}$ C	ECCON500BT	285/01	11.07.2028

Calibration Results

S.no	Standard Reading μ S/cm	UUC Reading(Mean) μ S/cm	Error μ S/cm
1	84	84.08	0.08
2	500	500.33	0.33
3	1413	1414.54	1.54

Remarks

- Results appeared are valid at the time of measurements.
- This certificate shall not be reproduced except in full without written approval of Euro
- Suggested calibration Due Date for Test Instrument : **1-Mar-2027**
- The above item is under non accredited parameter and it is traceable to national / international standards

** End of Certificate **

Checked By

Authorized Signatory

