



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, 11nd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 25

ULR Number : CC2519 26 0 00002830F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal- 637 214.

Date of Issue : 5-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0221

Receipt Date : 3-Mar-2026

Date of Calibration : 3-Mar-2026

Suggested calibration Due Date : 2-Mar-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-05

Unit Under Calibration

Name : Thermocouple With Thermometer
Make : Metravi
Model : DTM 900
S.No : 230902837
ID No. : IR-TC-226-04

Physical Condition : Good

Range : -200 to 1372 °C

Resolution : 0.1 / 1 °C

Accuracy : As per manual

Location : Chemical Lab

Temperature 25.6 °C

Environmental Conditions

Relative Humidity 60.9 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026
2	R-Type Thermocouple with 6 1/2 Digit Multimeter	1061& TW00022681	HT/CC/251213-20/002	12.12.2026

Calibration Results

S.No	Standard Reading °C	UUC Reading (Mean) °C	Error °C	Parameter - Thermal/Temperature	
				Error Allowed ±°C	Uncertainty ±°C
1	-29.95	-29.83	0.12	--	0.30
2	100.04	100.17	0.13	--	0.35
3	199.84	200.33	0.49	--	0.35
4	399.72	400.27	0.55	--	1.87
5	549.26	549.90	0.64	--	1.87
6	599.02	599.80	0.78	--	1.87
7	798.54	799.53	0.99	--	1.87
8	898.76	900.07	1.31	--	1.87

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT