



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 24

ULR Number : CC2519 26 0 00002829F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode ,
Namakkal- 637 214.

Date of Issue : 5-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0221

Receipt Date : 3-Mar-2026

Date of Calibration : 3-Mar-2026

Suggested calibration Due Date : 2-Mar-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-05

Unit Under Calibration

Physical Condition : Good

Name : Thermocouple With Thermometer

Make : Fisher Scientific

Model : Traceable

S.No : 101811135

ID No. : IR-TC-42-02

Range : -50 to 1300 °C

Resolution : 0.1 / 1 °C

Accuracy : As per manual

Location : Chemical Lab

Environmental Conditions

Temperature 25.5 °C

Relative Humidity 61.1 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026
2	R-Type Thermocouple with 6 1/2 Digit Multimeter	1061& TW00022681	HT/CC/251213-20/002	12.12.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	Standard Reading °C	UUC Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
1	-29.96	-29.87	0.09	--	0.30
2	99.64	99.83	0.19	--	0.35
3	199.23	199.33	0.10	--	0.67
4	399.54	400.33	0.79	--	1.94
5	549.33	550.00	0.67	--	1.87
6	599.12	600.00	0.88	--	1.87
7	799.28	801.33	2.05	--	1.94
8	898.87	901.33	2.46	--	1.94

** End of Certificate **



Checked By
(Hampiya .k.h)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT