



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 23

ULR Number : CC2519 26 0 00002744F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Unit Under Calibration

Physical Condition : Good

Name : BOD Incubator

Make : Servo

Model : Udian

S.No : 1473

ID No. : IR-IC-58-01

Range : 0 to 60 °C

Resolution : 0.1 °C

Accuracy : ± 1 °C

Location : Chemical Lab

Environmental Conditions

Temperature 26.5 °C

Relative Humidity 60 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018822 / 1711	ECT 2025/0971-01	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Temperature in °C		Standard Reading (Mean) °C	Error °C	Error Allowed $\pm$ °C	Uncertainty $\pm$ °C
	Set	Read				
1	25.0	25.0	25.27	-0.27	1.00	0.25
2	30.0	30.0	30.43	-0.43	1.00	0.25
3	40.0	40.0	40.63	-0.63	1.00	0.25
4	50.0	50.0	50.73	-0.73	1.00	0.36
5	55.0	55.0	55.87	-0.87	1.00	0.36
6	60.0	60.0	60.93	-0.93	1.00	0.36

** End of Certificate **



Checked By
(Haripriya.K.H)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

