



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IIInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 19

ULR Number : CC2519 26 0 00002740F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Shaking Water Bath (Temperature)

Make : Labtech / Selec

Model : LTWBIS-3

S.No : 18-C-101

ID No. : IR-SW-86-01

Range : Amb to 99 °C

Resolution : 1 °C

Accuracy : As Per Manual

Location : Chemical Lab

Environmental Conditions

Temperature : 25.8 °C

Relative Humidity : 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018822 / 1711	ECT 2025/0971-01	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	50	50	49.9	49.8	49.8	49.83	0.17	--	0.68
2	60	60	59.8	59.7	59.6	59.70	0.30	--	0.68
3	75	75	74.6	74.5	74.5	74.53	0.47	--	0.68
4	85	85	84.4	84.3	84.2	84.30	0.70	--	0.68
5	95	95	94.2	94.1	94.1	94.13	0.87	--	0.68

** End of Certificate **

Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH





EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 20

ULR Number : CC2519 26 0 00002741F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-09

Equipment Under Calibration

Physical Condition : Good

Name : Shaking Water Bath (RPM)

Make : Labtech / Selec

Model : LTWBIS-03

S.No : 18-C-101

ID No. : IR-SW-86-01

Range : 0 to 300 RPM

Resolution : 0.01 RPM

Accuracy : As Per Manual

Location : Chemical Lab

Environmental Conditions

Temperature : 25.5 °C

Relative Humidity : 58 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Tachometer	S622194	ECT 2025/1049-02	05.10.2026

Calibration Results

Parameter - Mechanical /Acceleration and Speed

S.No	UUC Reading RPM	Standard Reading in RPM			Standard Reading (Mean) RPM	Error RPM	Error Allowed ±RPM	Uncertainty ±RPM
		1	2	3				
1	50.00	49.7	49.7	49.8	49.73	0.27	--	3.80
2	100.00	99.7	99.6	99.5	99.60	0.40	--	3.80
3	150.00	149.3	149.2	149.3	149.27	0.73	--	3.80
4	200.00	199.1	199.0	198.9	199.00	1.00	--	3.80
5	295.00	293.3	293.4	293.2	293.30	1.70	--	3.80

** End of Certificate **



Remarks : See backside of this page

Checked By
(Signature)

Authorized Signatory
(Signature)
K. ABDUL VAJITH



ECT