



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IIInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 14

Date of Issue : 4-Mar-2026

ULR Number : CC2519 26 0 0002735F

Pages : 1 of 2

Customer Name and Address

SRF Number : 2026/0216

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 2-Mar-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 2-Mar-2026

Andipalayam Post, Tiruchengode,

Suggested calibration Due Date : 1-Mar-2027

Namakkal - 637 214.

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-05

Reference Standard : OIML R 76-1

Equipment Under Calibration

Physical Condition : Good

Name : Electronic Weighing Balance

Make : Point

Range : 1 g to 10 kg

Model : PTS-20

Resolution : 0.001 kg

S.No : 2015434

Accuracy : Class III

ID No. : RM-WB-108-04

Location : Raw Material Analysis Room
(Chemical Lab)

Environmental Conditions

Temperature 25.8 °C

Relative Humidity 59 %RH


Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Standard Weights	WI/09/2017/008-E1	TVCSPL 25/09/1032-01	01.09.2027
2	Standard Weights	501	HT/CC/250607-03/001	15.06.2026
3	Standard Weights	286, 738 & 739	RPL-2025-0727-C-007006	18.07.2026
4	Standard Weights	754	HT/CC/250726-03/001	27.07.2026

Parameter - Mechanical/Weighing Scale-Balance

1. Eccentricity

S.No.	Position of Weights used	UUC Reading (kg)	Error between centre(1) and Other points (kg)
1	3 (Top left)	2.000	0.000
2	4 (Top Right)	2.000	0.000
3	1 (Centre)	2.000	
4	2 (Bottom Left)	2.000	0.000
5	5 (Bottom Right)	2.000	0.000



Maximum Error between centre (1) and other points 0.000 kg



Checked By *(Hmipsiya.k.H)*

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 14
ULR Number : CC2519 26 0 00002735F

Date of Issue : 4-Mar-2026
Pages : 2 of 2

Calibration Results

Parameter - Mechanical/Weighing Scale-Balance

2. Error Test

S.No.	Standard Readings kg	UUC Readings kg	Error kg
1	0.01000	0.010	0.00000
2	0.02000	0.020	0.00000
3	0.05000	0.050	0.00000
4	0.10000	0.100	0.00000
5	0.20000	0.200	0.00000
6	0.50000	0.500	0.00000
7	1.00000	1.000	0.00000
8	2.00000	2.000	0.00000
9	4.99999	5.000	0.00001
10	10.00003	9.999	-0.00103

3.Repeatability

S.No.	UUC Reading			
	Zero (kg)	50% of range (kg)	Zero (kg)	100% of range (kg)
1	0.000	5.000	0.000	9.999
2	0.000	5.000	0.000	9.999
3	0.000	5.000	0.000	9.999
4	0.000	5.000	0.000	9.999
5	0.000	5.000	0.000	9.999
6	0.000	5.000	0.000	9.999
7	0.000	5.000	0.000	9.999
8	0.000	5.000	0.000	9.999
9	0.000	5.000	0.000	9.999
10	0.000	5.000	0.000	9.999

Expanded Uncertainty $\pm$ **0.00051** kg

Note : Uncertainty calculated for maximum capacity of balance for 10 repeated readings.

** End of Certificate **



Remarks : See backside of this page

(Handwritten signature)
Checked By

Authorized Signatory
K. ABDUL VAJITH



ECT