



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 23

ULR Number : CC2519 26 0 00002828F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 5-Mar-2026

Pages : 1 of 2

SRF Number : 2026/0221

Reciept Date : 3-Mar-2026

Date of Calibration : 3-Mar-2026

Suggested calibration Due Date : 2-Mar-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-04

Equipment Under Calibration

Physical Condition : Good

Name : Weight Box

Make : Wensar

Material : Stainless steel

S.No : 9318

ID No. : WL-WB-214

Range : *1 mg to 200 g

Resolution : NA

Accuracy : As Per Manual

Location : Chemical Lab

Environmental Conditions

Temperature 22.3 °C

Relative Humidity 58.1 %RH

Atmosphere Pressure

1008 mbar

{ Change in Temperature & Relative Humidity during calibration were less than ± 1.5 °C per hour and ± 15 % RH per 4 hours Respectively }

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Standard Weights	WI/09/2017/008-E1	TVCSP1 25/09/1032-01	01.09.2027

1. Calibration Method :

Comparison with Standards using Substitution Weighing(ABBA sequence)as per Standard OIML - R 111-1 : 2004

2. The Reported Mass Values are the Conventional Mass Value

3. Assumed Density (d) : (7950 $\pm$ 140) kg/m3: k=2

Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 23
ULR Number : CC2519 26 0 00002828F

Date of Issue : 5-Mar-2026
Pages : 2 of 2

Calibration Results

Parameter - Mechanical/Weights

S.no	Denomination mg	Standard Reading g	Error g	Uncertainty $\pm$ g
1	20	0.02000	0.00000	0.000014
*2	20	0.02002	-0.00002	0.000014
3	50	0.05002	-0.00002	0.000014
4	100	0.10001	-0.00001	0.000014
5	200	0.20001	-0.00001	0.000014
*6	200	0.20002	-0.00002	0.000014
7	500	0.50004	-0.00004	0.000014

S.no	Denomination g	Standard Reading g	Error g	Uncertainty $\pm$ g
8	1	1.00002	-0.00002	0.000014
9	2	1.99999	0.00001	0.000014
*10	2	2.00003	-0.00003	0.000014
11	5	5.00003	-0.00003	0.000014
12	10	10.00003	-0.00003	0.000020
13	20	19.99991	0.00009	0.000020
*14	20	20.00008	-0.00008	0.000020
15	50	50.00003	-0.00002	0.000020
16	100	100.00051	-0.00051	0.000139
17	200	200.00057	-0.00057	0.000140
*18	200	200.00047	-0.00047	0.000140

** End of Certificate **

(Hanipriya.KH)
Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT