



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 06

ULR Number : CC2519 26 0 00002811F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 5-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0221

Receipt Date : 3-Mar-2026

Date of Calibration : 3-Mar-2026

Suggested calibration Due Date : 2-Mar-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-03

Equipment Under Calibration

Physical Condition : Good

Name : Glass Thermometer

Make : Vision

Model : --

S.No : --

ID No. : CL-TM-215-01

Range : -10 to 110 °C

Resolution : 1 °C

Accuracy : As Per Manual

Location : Chemical Lab

Environmental Conditions

Temperature 25.3 °C

Relative Humidity 62.4 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	-10	-9.98	-9.98	-9.99	-9.98	-0.02	--	0.62
2	0	0.06	0.07	0.08	0.07	-0.07	--	0.62
3	10	10.11	10.12	10.11	10.11	-0.11	--	0.62
4	50	50.18	50.19	50.18	50.18	-0.18	--	0.62
5	110	110.40	110.41	110.42	110.41	-0.41	--	0.62

** End of Certificate **



Checked By
(Honipriya. K.H.)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT