



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 07
ULR Number : CC2519 26 0 00002812F
Customer Name and Address
M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 5-Mar-2026
Pages : 1 of 1
SRF Number : 2026/0221
Receipt Date : 3-Mar-2026
Date of Calibration : 3-Mar-2026
Suggested calibration Due Date : 2-Mar-2027
Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-03

Equipment Under Calibration

Physical Condition : Good

Name : Glass Thermometer
Make : Zeal
Model : --
S.No : --
ID No. : CL-GT-225-03

Range : -10 to 110 °C
Resolution : 1 °C
Accuracy : As Per Manual
Location : Chemical Lab

Environmental Conditions

Temperature 25.6 °C

Relative Humidity 61.5 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	-10	-10.07	-10.06	-10.06	-10.06	0.06	--	0.64
2	0	-0.09	-0.08	-0.08	-0.08	0.08	--	0.62
3	30	29.76	29.77	29.75	29.76	0.24	--	0.62
4	50	49.63	49.62	49.62	49.62	0.38	--	0.62
5	110	109.56	109.57	109.57	109.57	0.43	--	0.62

** End of Certificate **



Checked By
(Haripriya.KH)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT