



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 08
ULR Number : CC2519 26 0 00002813F
Customer Name and Address
M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 5-Mar-2026
Pages : 1 of 1
SRF Number : 2026/0221
Receipt Date : 3-Mar-2026
Date of Calibration : 3-Mar-2026
Suggested calibration Due Date : 2-Mar-2027
Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-03

Equipment Under Calibration

Physical Condition : Good

Name : Glass Thermometer
Make : SH
Jel : --
S.No : --
ID No. : CL-GT-255-03

Range : 0 to 250 °C
Resolution : 2 °C
Accuracy : As Per Manual
Location : Chemical Lab

Environmental Conditions

Temperature 25.5 °C

Relative Humidity 61.6 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	50	49.95	49.94	49.94	49.94	0.06	--	1.17
2	100	99.81	99.81	99.80	99.81	0.19	--	1.18
3	150	149.68	149.68	149.69	149.68	0.32	--	1.18
4	200	199.51	199.51	199.50	199.51	0.49	--	1.18
5	250	249.38	249.37	249.36	249.37	0.63	--	1.18

** End of Certificate **



Checked By
A. Hanipriya.KH

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT