



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 16
ULR Number : CC2519 26 0 00002737F
Customer Name and Address
M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 4-Mar-2026
Pages : 1 of 1
SRF Number : 2026/0216
Receipt Date : 2-Mar-2026
Date of Calibration : 2-Mar-2026
Suggested calibration Due Date : 1-Mar-2027
Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Name : Water Bath
Make : Guna / Selec
Model : TC544
S.No : 1391
ID No. : IR-WB-87-01

Physical Condition : Good
Range : Amb to 100 °C
Resolution : 1 °C
Accuracy : ± 2 °C
Location : Chemical Lab - Instrumentation Room
- IV

Environmental Conditions

Temperature 25.5 °C

Relative Humidity 59 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	50	50	50.1	50.2	50.1	50.13	-0.13	2.00	0.68
2	60	60	60.2	60.3	60.3	60.27	-0.27	2.00	0.68
3	80	80	80.5	80.6	80.6	80.57	-0.57	2.00	0.68
4	100	100	100.7	100.8	100.8	100.77	-0.77	2.00	0.68

** End of Certificate **



Checked By

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

