

EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 15

ULR Number : CC2519 26 0 00002736F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Rotary Evaporator Water Bath

Make : Equitron

Model : EV11

S.No : EV11.AEA.043

ID No. : IR-RW-91-01

Range : Amb+5 to 99 °C

Resolution : 1 °C

Accuracy : ± 1 °C

Location : Chemical Lab - Instrumentation Room
- III

Environmental Conditions

Temperature 25.7 °C

Relative Humidity 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	50	50	50.3	50.2	50.3	50.27	-0.27	1.00	0.68
2	60	60	60.4	60.5	60.5	60.47	-0.47	1.00	0.68
3	80	80	80.6	80.7	80.6	80.63	-0.63	1.00	0.68
4	95	95	95.7	95.7	95.8	95.73	-0.73	1.00	0.68

** End of Certificate **



(Haripriya.K.H)
Checked By

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT