



ECT/FT-17

EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 11
 ULR Number : CC2519 26 0 00002732F
 Customer Name and Address
 M/s. Christy Friedgram Industry Pvt Ltd,
 A2 & A3 SIDCO Industrial Estate,
 Andipalayam Post, Tiruchengode,
 Namakkal - 637 214.

Date of Issue : 4-Mar-2026
 Pages : 1 of 1
 SRF Number : 2026/0216
 Receipt Date : 2-Mar-2026
 Date of Calibration : 2-Mar-2026
 Suggested calibration Due Date : 1-Mar-2027
 Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerator With Freezer
 Make : LG
 Model : GL-308VT5/2009
 S.No : --
 ID No. : 36

Range : *2 to 8 °C / -18 to -20 °C
 Resolution : NA
 Accuracy : As Per Manual
 Location : Microbiology Lab

Environmental Conditions

Temperature 25.7 °C

Relative Humidity 57 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

Position	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ± °C
		1	2	3		
Rack 1	Refrigerator	3.7	3.8	3.9	3.80	0.68
Rack 2	2 to 8	3.5	3.6	3.5	3.53	
Rack 3		4.5	4.6	4.6	4.57	
Freezer	-18 to -20	-18.6	-18.6	-18.6	-18.60	

** End of Certificate **

Handwritten signature
 Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT