



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 10

ULR Number : CC2519 26 0 00002731F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerator With Freezer

Make : Godrej

Model : GDA19A1/2011

S.No : 110230735

ID No. : 35

Range : 2 to 8 °C / -18 to -20 °C

Resolution : NA

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature : 25.5 °C

Relative Humidity : 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

Position	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ±°C
		1	2	3		
Rack 1	Refrigerator	3.2	3.3	3.3	3.27	0.68
Rack 2	2 to 8	3.7	3.8	3.7	3.73	
Rack 3		5.1	5.2	5.3	5.20	
Freezer	-18 to -20	-18.3	-18.3	-18.3	-18.30	

** End of Certificate **



Checked By
(Signature)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

