



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 04

ULR Number : CC2519 25 0 00010508F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 12-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0966

Receipt Date : 10-Sep-2025

Date of Calibration : 10-Sep-2025

Suggested calibration Due Date : 9-Sep-2026

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerator With Freezer

Make : Samsung

Model : RT28T31429R/HL/2020

S.No : --

ID No. : 34

Range : 2 to 8 °C / -18 to -20 °C

Resolution : NA

Accuracy : As per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature : 26.4 °C

Relative Humidity : 57 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Calibrator with Sensor	091109171/1703	ECT 2025/0142-01	04.02.2026

Calibration Results

Parameter - Thermal / Temperature

Position	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ±°C
		1	2	3		
	Refrigerator					
Rack 1	2 to 8	6.4	6.3	6.3	6.33	0.90
Rack 2	2 to 8	6.3	6.4	6.3	6.33	0.90
Rack 3	2 to 8	6.2	6.1	6.1	6.13	0.90
	Freezer					
Rack 1	-18 to -20	-18.2	-18.3	-18.4	-18.30	0.91

** End of Certificate **



Checked By
Hampiya.K.H.

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



2025