



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 02
ULR Number : CC2519 25 0 00010506F
Customer Name and Address
M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode ,
Namakkal- 637 214.

Date of Issue : 12-Sep-2025
Pages : 1 of 1
SRF Number : 2025/0966
Receipt Date : 10-Sep-2025
Date of Calibration : 10-Sep-2025
Suggested calibration Due Date : 9-Sep-2026
Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Name : Incubator-1
Make : Servo Enterprises
Model : --
S.No : 1474
ID No. : 19

Physical Condition : Good
Range : Amb to 60 °C
Resolution : 0.1 °C
Accuracy : ± 1 °C
Location : Incubation Room - Microbiology Lab

Environmental Conditions

Temperature : 25.8 °C Relative Humidity : 59 %RH Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Calibrator with Sensor	091109171/1704	ECT 2025/0142-02	04.02.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	30.0	30.0	30.1	30.1	30.1	30.10	-0.10	1.00	0.35
2	35.0	35.0	35.2	35.1	35.1	35.13	-0.13	1.00	0.36
3	37.0	37.0	37.2	37.2	37.3	37.23	-0.23	1.00	0.36
4	42.0	42.0	42.4	42.4	42.5	42.43	-0.43	1.00	0.36
5	44.0	44.0	44.6	44.6	44.6	44.60	-0.60	1.00	0.35
6	45.0	45.0	45.6	45.6	45.7	45.63	-0.63	1.00	0.36

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

