



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2025/0966- 03

ULR Number : CC2519 25 0 00010507F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 12-Sep-2025

Pages : 1 of 1

SRF Number : 2025/0966

Receipt Date : 10-Sep-2025

Date of Calibration : 10-Sep-2025

Suggested calibration Due Date : 9-Sep-2026

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Incubator-2

Make : Servo Enterprises

Model : --

S.No : 1474

ID No. : 20

Range : Amb to 60 °C

Resolution : 0.1 °C

Accuracy : ± 1 °C

Location : Incubation Room - Microbiology Lab

Environmental Conditions

Temperature 25.9 °C

Relative Humidity 59 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Calibrator with Sensor	091109171/1705	ECT 2025/0142-03	04.02.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	30.0	30.0	30.1	30.2	30.2	30.17	-0.17	1.00	0.36
2	35.0	35.0	35.2	35.3	35.3	35.27	-0.27	1.00	0.36
3	37.0	37.0	37.4	37.4	37.4	37.40	-0.40	1.00	0.35
4	42.0	42.0	42.5	42.5	42.5	42.50	-0.50	1.00	0.35
5	44.0	44.0	44.6	44.5	44.5	44.53	-0.53	1.00	0.36
6	55.0	55.0	55.7	55.7	55.6	55.67	-0.67	1.00	0.36

** End of Certificate **

(Haripriya.K.H.)
Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

