



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 02
ULR Number : CC2519 26 0 00005768F
Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode ,
Namakkal- 637 214.

Date of Issue : 9-Jun-2026
Pages : 1 of 1
SRF Number : 2026/0493
Receipt Date : 8-Jun-2026
Date of Calibration : 8-Jun-2026
Suggested calibration Due Date : 7-Jun-2027
Calibrated at : Onsite

Customer Reference : Work Order

Unit Under Calibration

Calibration Procedure : ECT/WI-T-05

Physical Condition : Good

Name : Digital Temperature controller with sensor (Autoclave)
Make : Pisces/Selec
Model : TC 533
S.No : A-147
ID No. : 09

Range : Amb to 125 °C
Resolution : 0.1 °C
Accuracy : ± 3 °C
Location : Microbiology Lab/Sterilization Room

Environmental Conditions
Temperature 26.4 °C Relative Humidity 59 %RH Temperature Scale : ITS 90
Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018822 / 1711	ECT 2025/0971-01	10.09.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	Standard Reading °C	UUC Reading in °C			UUC Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	109.8	110.6	110.5	110.6	110.57	0.77	3.00	0.36
2	114.6	115.9	115.9	115.8	115.87	1.27	3.00	0.36
3	120.4	121.8	121.8	121.7	121.77	1.37	3.00	0.36


** End of Certificate **

Checked By
(Haripriya.K.H.)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH





EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 03

ULR Number : CC2519 26 0 00005769F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode ,
Namakkal- 637 214.

Date of Issue : 9-Jun-2026

Pages : 1 of 1

SRF Number : 2026/0493

Receipt Date : 8-Jun-2026

Date of Calibration : 8-Jun-2026

Suggested calibration Due Date : 7-Jun-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-01

Unit Under Calibration

Physical Condition : Good

Name : Pressure Gauge (Autoclave)

Make : Kamal

Model : --

S.No : --

ID No. : 09

Range : 0 to 60 lb/in²

Resolution : 1 lb/in²

Accuracy : As per Manual

Location : Microbiology Lab/Sterilization Room

Environmental Conditions

Temperature 26.4 °C

Relative Humidity 59 %RH

Lab- Temperature 25 ± 2°C / Humidity 30 to 75 %RH Onsite Temperature 25 ± 5 °C / Humidity 30 to 75 %RH

Change in Temperature during Calibration not exceeded ± 1 °C

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Pressure Gauge	3108774	WCP/025-720,WCP/025-721	28.07.2026

Calibration Results

Parameter - Mechanical/Pressure Indicating Devices

S.No	UUC Instrument	Standard Reading		Mean (X)	Hysteresis (h)	Deviation ΔP	Error Allowed	Expanded Uncertainty
		UP	DOWN					
	lb/in ²	lb/in ²	lb/in ²	lb/in ²	lb/in ²	lb/in ²	± lb/in ²	± lb/in ²
1	0	0.00	0.00	0.000	0.000	0.000	--	--
2	10	10.14	10.16	10.150	0.020	-0.150	--	0.578
3	15	15.22	15.24	15.230	0.020	-0.230	--	0.578
4	40	40.44	40.52	40.480	0.080	-0.480	--	0.585
5	60	60.74	60.74	60.740	0.000	-0.740	--	0.577

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH





ECT/FT-17

EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IIInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 04

ULR Number : CC2519 26 0 00005770F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 13-Jun-2026

Pages : 1 of 1

SRF Number : 2026/0493

Receipt Date : 8-Jun-2026

Date of Calibration : 8-Jun-2026

Suggested calibration Due Date : 7-Jun-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-E-01

Unit Under Calibration

Physical Condition : Good

Name : Timer (Autoclave)

Make : Selec

Model : TC 533

S.No : --

ID No. : 09

Range : 0 to 30 min

Resolution : 1 min

Accuracy : As per Manual

Location : Microbiology Lab/Sterilization Room

Environmental Conditions

Temperature 26.4 °C

Relative Humidity 59 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	ID No	Certificate number	Validity
1	Digital Timer	ECT/E/TIM/06	ECT 2026/0455-01	20.05.2027

Calibration Results

Parameter - Electro Technical

S.no	UUC Reading Min	Standard Reading				Error				Uncertainty ± sec
		Hr	Min	Sec	1/100s	Hr	Min	Sec	1/100s	
1	10	0	10	0	9	0	0	0	-9	1.33
2	15	0	15	0	19	0	0	0	-19	1.33
3	20	0	20	0	28	0	0	0	-28	1.34
4	30	0	30	0	33	0	0	0	-33	1.34


** End of Certificate **

(Haripriya.K.H)
Checked By



Remarks : See backside of this page

(Signature)
Authorized Signatory
K. ABDUL VAJITH



2027