



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 09

ULR Number : CC2519 26 0 00002730F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Incubator

Make : Servo Scientific Suppliers

Model : --

S.No : 20251104

ID No. : 7

Range : Amb to 100 °C

Resolution : 0.1 °C

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature : 25.4 °C

Relative Humidity : 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
			1	2	3				
1	40.0	40.0	39.8	39.9	39.8	39.83	0.17	--	0.36
2	42.0	42.0	41.7	41.7	41.6	41.67	0.33	--	0.36
3	45.0	45.0	44.5	44.6	44.5	44.53	0.47	--	0.36
4	55.0	55.0	54.4	54.3	54.3	54.33	0.67	--	0.36
5	65.0	65.0	64.3	64.3	64.2	64.27	0.73	--	0.36
6	80.0	80.0	79.2	79.1	79.2	79.17	0.83	--	0.36

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT