

EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0294- 02

ULR Number : CC2519 26 0 00003880F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt. Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode

Namakkal - 637 214.

Date of Issue : 23-Mar-2026

Pages : 1 of 2

SRF Number : 2026/0294

Receipt Date : 20-Mar-2026

Date of Calibration : 21-Mar-2026

Suggested calibration Due Date : 20-Mar-2027

Calibrated at In House

Calibration Procedure : ECT/WI-M-06-1

Reference Standard : ISO 8655-6

Customer Reference : Work Order

Equipment Under Calibration

Physical Condition : Good

Name : Micropipette

Make : Eppendorf

Model : --

S.No : Q51644J

ID No. : 05

Range : 10 to 100 μ l

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature 22.1 $^{\circ}$ C

Relative Humidity 57.9 %RH

Atmosphere Pressure

1008 mbar

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Electronic Semi Micro Balance	31504484	HT/CC/251213-31/001	14.12.2026

Calibration Results

Parameter - Mechanical / Volume

1. Volume Measurement for Low Range 20 μ l

Number of Measurements taken is 10

Measurments									
μ l									
1	2	3	4	5	6	7	8	9	10
19.65	19.89	19.98	19.72	19.68	19.86	19.78	19.89	19.62	19.82

Mean Volume 19.79 μ l

Systematic Error -0.21 μ l

Random Error 0.12 μ l


Measurement Uncertainty $\pm$ 0.38 μ l


Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0294- 02

ULR Number : CC2519 26 0 00003880F

Date of Issue : 23-Mar-2026

Pages : 2 of 2

2. Volume Measurement for Middle Range 50 μ l
Number of Measurements taken is 10

Measurments									
μ l									
1	2	3	4	5	6	7	8	9	10
49.96	49.98	50.02	50.32	50.18	49.96	49.87	49.92	49.68	49.98

Mean Volume 49.99 μ l
Systematic Error -0.01 μ l
Random Error 0.17 μ l
Measurement Uncertainty $\pm$ 0.38 μ l

3. Volume Measurement for High Range 100 μ l
Number of Measurements taken is 10

Measurments									
μ l									
1	2	3	4	5	6	7	8	9	10
99.96	100.02	100.35	100.21	99.85	99.65	100.09	100.25	100.24	100.09

Mean Volume 100.07 μ l
Systematic Error 0.07 μ l
Random Error 0.21 μ l
Measurement Uncertainty $\pm$ 0.38 μ l

** End of Certificate **

Checked By
(Hampi...)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

