



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 03

ULR Number : CC2519 26 0 00002724F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Incubator

Make : Heco

Model : TMS02

S.No : 337

ID No. : 06

Range : Amb to 100 °C

Resolution : 0.1 °C

Accuracy : ± 1 °C

Location : Microbiology (Incubation Room)

Environmental Conditions

Temperature 25.6 °C

Relative Humidity 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed $\pm$ °C	Uncertainty $\pm$ °C
			1	2	3				
1	40.0	40.0	40.1	40.2	40.1	40.13	-0.13	1.00	0.36
2	42.0	42.0	42.2	42.3	42.3	42.27	-0.27	1.00	0.36
3	45.0	45.0	45.3	45.3	45.3	45.30	-0.30	1.00	0.35
4	55.0	55.0	55.3	55.4	55.5	55.40	-0.40	1.00	0.37
5	65.0	65.0	65.6	65.6	65.7	65.63	-0.63	1.00	0.36
6	80.0	80.0	80.7	80.8	80.8	80.77	-0.77	1.00	0.36

** End of Certificate **


Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

