




EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 04

ULR Number : CC2519 26 0 00002725F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Water Bath

Make : Kemi

Model : KWB-220

S.No : 12201

ID No. : 8

Range : 30 to 110 °C

Resolution : NA

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature : 25.3 °C

Relative Humidity : 57 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	40	40.3	40.4	40.3	40.33	-0.33	--	0.36
2	60	60.6	60.5	60.5	60.53	-0.53	--	0.36
3	80	80.7	80.7	80.6	80.67	-0.67	--	0.36
4	100	100.8	100.9	100.8	100.83	-0.83	--	0.36

** End of Certificate **

Checked By
H. Haripriya.KH



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT