



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 22

ULR Number : CC2519 26 0 00002827F

Date of Issue : 5-Mar-2026

Pages : 2 of 2

Calibration Results

Parameter - Mechanical/Weights

S.no	Denomination mg	Standard Reading g	Error g	Maximum Permissible limit ± g	Uncertainty ± g
1	20	0.01998	0.00003	0.00030	0.000014
*2	20	0.01997	0.00003	0.00030	0.000014
3	50	0.05001	-0.00001	0.00040	0.000014
4	100	0.09996	0.00004	0.00050	0.000014
5	200	0.20002	-0.00002	0.00060	0.000014
*6	200	0.20004	-0.00003	0.00060	0.000014
7	500	0.49997	0.00003	0.00080	0.000014

S.no	Denomination g	Standard Reading g	Error g	Maximum Permissible limit ± g	Uncertainty ± g
8	1	1.00003	-0.00003	0.00100	0.000014
9	2	2.00006	-0.00006	0.00120	0.000014
*10	2	1.99995	0.00005	0.00120	0.000014
11	5	4.99997	0.00003	0.00160	0.000014
12	10	10.00005	-0.00005	0.00200	0.000020
13	20	20.00004	-0.00004	0.00250	0.000020
*14	20	20.00008	-0.00008	0.00250	0.000020
15	50	49.99993	0.00008	0.00300	0.000020
16	100	100.00031	-0.00030	0.0050	0.000139
17	200	200.00052	-0.00052	0.0100	0.000140

Note : Mass Value of all the weights are within the maximum errors permissible in M1 accuracy class of weights as per OIML Recommendation No R111-1 (2004)

** End of Certificate **


Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



8/17



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 22

ULR Number : CC2519 26 0 00002827F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 5-Mar-2026

Pages : 1 of 2

SRF Number : 2026/0221

Reciept Date : 3-Mar-2026

Date of Calibration : 3-Mar-2026

Suggested calibration Due Date : 2-Mar-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-04

Equipment Under Calibration

Physical Condition : Good

Name : Standard Weight Box

Make : Kinglab

Material : --

S.No : 1214

ID No. : 02

Range : 1 mg to 200 g

Resolution : NA

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature 22.3 °C

Relative Humidity 57.9 %RH

Atmosphere Pressure 1008 mbar

{ Change in Temperature & Relative Humidity during calibration were less than ± 1.5 °C per hour and ± 15 % RH per 4 hours
Respectively }

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Standard Weights	WI/09/2017/008-E1	TVCSPL 25/09/1032-01	01.09.2027

1. Calibration Method :

Comparison with Standards using Substitution Weighing (ABBA sequence) as per Standard OIML - R 111-1 : 2004

2. The Reported Mass Values are the Conventional Mass Value

3. Assumed Density (d) : (7950 $\pm$ 140) kg/m³: k=2



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



Checked By
(Hamipriya K.H.)