



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0294- 01

ULR Number : CC2519 26 0 00003879F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt. Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode

Namakkal - 637 214.

Date of Issue : 23-Mar-2026

Pages : 1 of 2

SRF Number : 2026/0294

Receipt Date : 20-Mar-2026

Date of Calibration : 21-Mar-2026

Suggested calibration Due Date : 20-Mar-2027

Calibrated at In House

Calibration Procedure : ECT/WI-M-06-1

Reference Standard : ISO 8655-6

Customer Reference : Work Order

Equipment Under Calibration

Physical Condition : Good

Name : Micropipette

Make : Eppendorf

Model : Research Plus

S.No : R16833E

ID No. : 03

Range : 100 to 1000 μ l

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature 22.4 $^{\circ}$ C

Relative Humidity 57.8 %RH Atmosphere Pressure

1008 mbar

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Electronic Semi Micro Balance	31504484	HT/CC/251213-31/001	14.12.2026

Calibration Results

Parameter - Mechanical / Volume

1. Volume Measurement for Low Range 100 μ l

Number of Measurements taken is 10

Measurements									
μ l									
1	2	3	4	5	6	7	8	9	10
99.58	99.64	99.51	99.40	99.38	99.73	99.81	99.90	99.68	99.79

Mean Volume 99.64 μ l

Systematic Error -0.36 μ l

Random Error 0.17 μ l

Measurement Uncertainty $\pm$ 1.8 μ l


Checked By



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH






EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, 11nd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0294- 01

ULR Number : CC2519 26 0 00003879F

Date of Issue : 23-Mar-2026

Pages : 2 of 2

2. Volume Measurement for Middle Range 500 μ l
Number of Measurements taken is 10

Measurments									
μ l									
1	2	3	4	5	6	7	8	9	10
499.85	499.76	499.89	499.40	499.28	499.56	499.24	499.36	499.70	499.65

Mean Volume 499.57 μ l
Systematic Error -0.43 μ l
Random Error 0.24 μ l
Measurement Uncertainty $\pm$ 1.8 μ l

3. Volume Measurement for High Range 1000 μ l
Number of Measurements taken is 10

Measurments									
μ l									
1	2	3	4	5	6	7	8	9	10
999.29	999.40	999.36	999.41	999.50	999.40	999.48	999.53	999.64	999.34

Mean Volume 999.44 μ l
Systematic Error -0.56 μ l
Random Error 0.10 μ l
Measurement Uncertainty $\pm$ 1.8 μ l

** End of Certificate **

Checked By
(Haripriya K.N.)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

