



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 05

ULR Number : CC2519 26 0 00002726F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-05

Equipment Under Calibration

Physical Condition : Good

Name : Digital Temperature Controller With Sensor (Autoclave)

Make : Servo Scientific Suppliers

Model : --

S.No : 20251216

ID No. : 10

Range : Amb to 121 °C

Resolution : 0.1 °C

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature : 25.5 °C

Relative Humidity : 58 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018808 / 1717	ECT 2025/0971-04	10.09.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	Standard Reading °C	UUC Reading in °C			UUC Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	110.0	109.9	109.8	109.8	109.83	-0.17	--	0.36
2	115.0	114.7	114.8	114.8	114.77	-0.23	--	0.36
2	121.0	120.6	120.7	120.5	120.60	-0.40	--	0.37

** End of Certificate **

Checked By : (Hemipriya. K.H.)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH





EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 06

ULR Number : CC2519 26 0 00002727F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-01

Equipment Under Calibration

Physical Condition : Good

Name : Pressure Gauge

Make : Baumer

Model : EN837-1

S.No : --

ID No. : 10

Range : 0 to 60 Psi

Resolution : 1 Psi

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature 25.4 °C

Relative Humidity 58 %RH

Lab- Temperature 25 ± 2°C / Humidity 30 to 75 %RH Onsite Temperature 25 ± 5 °C / Humidity 30 to 75 %RH

Change in Temperature during Calibration not exceeded ± 1 °C

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Pressure Gauge	3108774	WCP/025-720,WCP/025-721	28.07.2026

Calibration Results

Parameter - Mechanical/Pressure Indicating Devices

S.No	UUC Instrument	Standard Reading		Mean (X)	Hysteresis (h)	Deviation ΔP	Error Allowed	Expanded Uncertainty
		UP	DOWN					
	Psi	Psi	Psi	Psi	Psi	Psi	± Psi	± Psi
1	0	0.00	0.00	0.000	0.000	0.000	--	--
2	5	5.12	5.09	5.105	-0.030	-0.105	--	0.578
3	15	15.18	15.15	15.165	-0.030	-0.165	--	0.578
4	20	20.25	20.21	20.230	-0.040	-0.230	--	0.579
5	30	30.36	30.30	30.330	-0.060	-0.330	--	0.582
6	50	50.46	50.46	50.460	0.000	-0.460	--	0.577

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT

Checked By
(Hampiya. KH)



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IIInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 07

ULR Number : CC2519 26 0 00002728F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 4-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0216

Receipt Date : 2-Mar-2026

Date of Calibration : 2-Mar-2026

Suggested calibration Due Date : 1-Mar-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-ET-01

Equipment Under Calibration

Physical Condition : Good

Name : Timer (Autoclave)
Make : Servo Scientific Suppliers
Model : --
S.No : 20251216
ID No. : 10

Range : 0 to 30 min
Resolution : 1 min
Accuracy : As Per Manual
Location : Microbiology Lab

Environmental Conditions

Temperature : 25.4 °C

Relative Humidity : 58 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	ID No	Certificate number	Validity
1	Digital Timer	ECT/E/TIM/06	ECT 2025/0567-01	21.05.2026

Calibration Results

Parameter - Electro Technical

S.no	UUC Reading Min	Eq. Standard Reading				Error				Uncertainty ± Sec
		Hr	Min	Sec	1/100s	Hr	Min	Sec	1/100s	
1	5	0	5	0	7	00	00	00	-7	1.2
2	10	0	10	0	13	00	00	00	-13	
3	15	0	15	0	18	00	00	00	-18	
4	20	0	20	0	26	00	00	00	-26	
5	30	0	30	0	32	00	00	00	-32	

** End of Certificate **



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



(Hampiya.K.H.)
Checked By