



ECT/FT-17

EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 06

Date of Issue : 9-Jun-2026

ULR Number : CC2519 26 0 00005772F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2026/0493

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 8-Jun-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 8-Jun-2026

Andipalayam Post, Tiruchengode ,

Suggested calibration Due Date : 7-Jun-2027

Namakkal- 637 214.

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-05

Equipment Under Calibration

Physical Condition : Good

Name : Digital Temperature controller with sensor (Autoclave)

Make : Kemi

Range : Amb to 125 °C

Model : KAUC-S-5

Resolution : 0.1 °C

S.No : 2051

Accuracy : ± 3 °C

ID No. : 14

Location :

Microbiology Lab/Sterilization Room

Environmental Conditions

Temperature 26.6 °C

Relative Humidity 61 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018822 / 1711	ECT 2025/0971-01	10.09.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	Standard Reading °C	UUC Reading in °C			UUC Reading (Mean) °C	Error °C	Error Allowed $\pm$ °C	Uncertainty $\pm$ °C
		1	2	3				
1	110.4	110.6	110.5	110.6	110.57	0.17	3.00	0.36
2	115.7	115.9	115.9	116.0	115.93	0.23	3.00	0.36
3	122.3	122.6	122.6	122.7	122.63	0.33	3.00	0.36

** End of Certificate **



Checked By
H. Haripriya (K.H.)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



2027



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 08

ULR Number : CC2519 26 0 00005774F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 9-Jun-2026

Pages : 1 of 1

SRF Number : 2026/0493

Receipt Date : 8-Jun-2026

Date of Calibration : 8-Jun-2026

Suggested calibration Due Date : 7-Jun-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-01

Unit Under Calibration

Physical Condition : Good

Name : Pressure Gauge (Autoclave)

Make : Accurate

Model : --

S.No : --

ID No. : 14

Range : 0 to 30 psi

Resolution : 1 psi

Accuracy : As per Manual

Location : Microbiology Lab/Sterilization Room

Environmental Conditions

Temperature 26.5 °C

Relative Humidity 60 %RH

Lab- Temperature 25 ± 2°C / Humidity 30 to 75 %RH Onsite Temperature 25 ± 5 °C / Humidity 30 to 75 %RH

Change in Temperature during Calibration not exceeded ± 1 °C

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Pressure Gauge	3108774	WCP/025-720,WCP/025-721	28.07.2026

Calibration Results

Parameter - Mechanical/Pressure Indicating Devices

S.No	UUC Instrument	Standard Reading		Mean (X)	Hysteresis (h)	Deviation ΔP	Error Allowed	Expanded Uncertainty
		UP	DOWN					
	psi	psi	psi	psi	psi	psi	± psi	± psi
1	0	0.00	0.00	0.000	0.000	0.000	--	--
2	10	9.92	9.84	9.880	-0.080	0.120	--	0.585
3	15	14.83	14.76	14.795	-0.070	0.205	--	0.583
4	20	19.65	19.54	19.595	-0.110	0.405	--	0.591
5	30	29.47	29.47	29.470	0.000	0.530	--	0.577

** End of Certificate **



Checked By
(Hemipriya.K.H)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 10

ULR Number : CC2519 26 0 00005776F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 13-Jun-2026

Pages : 1 of 1

SRF Number : 2026/0493

Receipt Date : 8-Jun-2026

Date of Calibration : 8-Jun-2026

Suggested calibration Due Date : 7-Jun-2027

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-E-01

Unit Under Calibration

Physical Condition : Good

Name : Timer (Autoclave)

Make : Kemi

Model : --

S.No : --

ID No. : 14

Range : 0 to 30 min

Resolution : 1 min

Accuracy : As per Manual

Location : Microbiology Lab/Sterilization Room

Environmental Conditions

Temperature 26.8 °C

Relative Humidity 61 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	ID No	Certificate number	Validity
1	Digital Timer	ECT/E/TIM/06	ECT 2026/0455-01	20.05.2027

Calibration Results

Parameter - Electro Technical

S.no	UUC Reading Min	Standard Reading				Error				Uncertainty ± sec
		Hr	Min	Sec	1/100s	Hr	Min	Sec	1/100s	
1	10	0	10	0	16	0	0	0	-16	1.33
2	15	0	15	0	23	0	0	0	-23	1.33
3	20	0	20	0	44	0	0	0	-44	1.33
4	30	0	30	0	32	0	0	0	-32	1.33

** End of Certificate **

Checked By *(Hampiya K.H)*



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH



ECT