



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 05

ULR Number : CC2519 26 0 00005771F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode ,

Namakkal- 637 214.

Date of Issue : 9-Jun-2026

Pages : 1 of 1

SRF Number : 2026/0493

Receipt Date : 8-Jun-2026

Date of Calibration : 8-Jun-2026

Suggested calibration Due Date : 7-Jun-2027

Calibrated at Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-05

Equipment Under Calibration

Name : Digital Temperature controller with sensor (Autoclave)

Make : Labtech/Selec

Model : TC 544

S.No : --

ID No. : 11

Range : Amb to 125 °C

Resolution : 0.1 °C

Accuracy : ± 3 °C

Location : Microbiology Lab/Decontamination Room

Physical Condition : Good

Environmental Conditions

Temperature 26.5 °C

Relative Humidity 61 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature Calibrator With RTD Sensor	220615018822 / 1711	ECT 2025/0971-01	10.09.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	Standard Reading °C	UUC Reading in °C			UUC Reading (Mean) °C	Error °C	Error Allowed $\pm$ °C	Uncertainty $\pm$ °C
		1	2	3				
1	110.8	110.6	110.7	110.6	110.63	-0.17	3.00	0.36
2	115.6	115.4	115.3	115.4	115.37	-0.23	3.00	0.36
3	121.4	121.1	121.1	121.1	121.10	-0.30	3.00	0.35

** End of Certificate **

Checked By
(Haripriya K.H.)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH





EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CC-2519

CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 07

Date of Issue : 9-Jun-2026

ULR Number : CC2519 26 0 00005773F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2026/0493

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 8-Jun-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 8-Jun-2026

Andipalayam Post, Tiruchengode ,

Suggested calibration Due Date : 7-Jun-2027

Namakkal- 637 214.

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-01

Unit Under Calibration

Physical Condition : Good

Name : Pressure Gauge (Autoclave)

Make : Accurate

Range : 0 to 30 psi

Model : --

Resolution : 1 psi

S.No : --

Accuracy : As per Manual

ID No. : 11

Location : Microbiology Lab / Decontamination Room

Environmental Conditions

Temperature 26.5 °C

Relative Humidity 60 %RH

Lab- Temperature 25 ± 2°C / Humidity 30 to 75 %RH Onsite Temperature 25 ± 5 °C / Humidity 30 to 75 %RH

Change in Temperature during Calibration not exceeded ± 1 °C

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Pressure Gauge	3108774	WCP/025-720,WCP/025-721	28.07.2026

Calibration Results

Parameter - Mechanical/Pressure Indicating Devices

S.No	UUC Instrument	Standard Reading		Mean (X)	Hysteresis (h)	Deviation ΔP	Error Allowed	Expanded Uncertainty
		UP	DOWN					
	psi	psi	psi	psi	psi	psi	± psi	± psi
1	0	0.00	0.00	0.000	0.000	0.000	--	--
2	10	9.84	9.86	9.850	0.020	0.150	--	0.578
3	15	14.74	14.78	14.760	0.040	0.240	--	0.579
4	20	19.66	19.82	19.740	0.160	0.260	--	0.606
5	30	29.58	29.58	29.580	0.000	0.420	--	0.577

** End of Certificate **



Checked By
B. Hampriya . K.H

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH





EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0493- 09

Date of Issue : 13-Jun-2026

ULR Number : CC2519 26 0 00005775F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2026/0493

M/s. Christy Friedgram Industry Pvt-Ltd,

Receipt Date : 8-Jun-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 8-Jun-2026

Andipalayam Post, Tiruchengode ,

Suggested calibration Due Date : 7-Jun-2027

Namakkal- 637 214.

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-E-01

Unit Under Calibration

Physical Condition : Good

Name : Timer (Autoclave)

Make : Selec

Range : 0 to 30 min

Model : TC 544

Resolution : 1 min

S.No : --

Accuracy : As per Manual

ID No. : 11

Location : Microbiology Lab /
Decontamination Room

Environmental Conditions

Temperature 26.6 °C

Relative Humidity 61 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	ID No	Certificate number	Validity
1	Digital Timer	ECT/E/TIM/06	ECT 2026/0455-01	20.05.2027

Calibration Results

Parameter - Electro Technical

S.no	UUC Reading Min	Standard Reading				Error				Uncertainty ± sec
		Hr	Min	Sec	1/100s	Hr	Min	Sec	1/100s	
1	10	0	10	0	16	0	0	0	-16	1.33
2	15	0	15	0	26	0	0	0	-26	1.33
3	20	0	20	0	33	0	0	0	-33	1.33
4	30	0	30	0	45	0	0	0	-45	1.33

** End of Certificate **



Remarks : See backside of this page

Checked By
(Haripriya.k.h)

Authorized Signatory
K. ABDUL VAJITH

