



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, € Certificate Number : ECT/ NN 2026/0110 -02
Ambattur, Chennai - 600 053. Tamil Nadu. € Receipt Number : NN 2026/0110 / 2-Mar-2026
Tel : 044 - 4852 1288. € Calibration Date : 2-Mar-2026
Mobile : 9486195184, 9176032184. € Date of Issue : 4-Mar-2026 Page 01 of 01

CERTIFICATE OF CALIBRATION

Name and Address of Customer

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode
Namakkal - 637 214.

Customer Reference

Work order

Calibrated at : Onsite

Equipment Under Calibration

Name : pH Meter
Make : Eutech Instruments
Model : PH700
S.No : 3238268
ID No : 12

Range : 0 to 14 pH
Resolution : 0.01 pH
Accuracy : As Per Manual
Location : Microbiology Lab

Environmental Condition :

Temperature : 25.5 °C
Humidity : 58 % RH

Calibration Procedure : ECT/WI-NN-01

Physical condition : Good

Standard Equipment used for Calibration

S.No	Description	Product Code	Batch No	Master Validity
1	4.01pH @25°C pH Buffer	ECBU4BT	424C1	28.03.2027
2	7.00pH @25°C pH Buffer	ECBU7BT	72524C1	28.03.2026
3	10.01pH @25°C pH Buffer	ECBU10BT	375/02	12.09.2028

Calibration Results

S.no	Standard Reading PH	UUC Reading(Mean) PH	Error PH
1	4.01	4.04	0.03
2	7.00	7.01	0.01
3	10.01	10.02	0.01

Remarks

- Results appeared are valid at the time of measurements.
- This certificate shall not be reproduced except in full without written approval of Euro
- Suggested calibration Due Date for Test Instrument : **1-Mar-2027**
- The above item is under non accredited parameter and it is traceable to national / international standards

** End of Certificate **

Checked By

Authorized Signatory

