



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0216- 02

Date of Issue : 4-Mar-2026

ULR Number : CC2519 26 0 00002723F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2026/0216

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 2-Mar-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 2-Mar-2026

Andipalayam Post, Tiruchengode,

Suggested calibration Due Date : 1-Mar-2027

Namakkal - 637 214.

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerated Circulating Water Bath

Make : Labman

Range : -20 to 100 °C

Model : RCB1220

Resolution : 0.01 °C

S.No : 137874

Accuracy : ± 0.5 °C

ID No. : 13

Location : Microbiology Lab (Incubation Room)

Environmental Conditions

Temperature : 25.4 °C

Relative Humidity : 57 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Thermometer With RTD Sensor	22000243 &T0256/1719	SAN/25/1042/TH/1	11.08.2026

Calibration Results

Parameter - Thermal / Temperature

S.No	UUC Set Temperature °C	UUC Read Temperature °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed $\pm$ °C	Uncertainty $\pm$ °C
			1	2	3				
1	8.00	8.00	8.1	8.0	8.1	8.07	-0.07	0.50	0.220
2	15.00	15.00	15.1	15.2	15.2	15.17	-0.17	0.50	0.220
3	20.00	20.00	20.2	20.2	20.2	20.20	-0.20	0.50	0.220
4	35.00	35.00	35.2	35.3	35.3	35.27	-0.27	0.50	0.220
5	45.00	45.00	45.3	45.3	45.4	45.33	-0.33	0.50	0.220
6	60.00	60.00	60.5	60.6	60.7	60.60	-0.60	0.50	0.350
7	80.00	80.00	80.7	80.7	80.8	80.73	-0.73	0.50	0.350

** End of Certificate **



Checked By : H. H. Priya . KH

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

