



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 02
ULR Number : CC2519 26 0 00002807F
Customer Name and Address
M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 5-Mar-2026
Pages : 1 of 1
SRF Number : 2026/0221
Receipt Date : 3-Mar-2026
Date of Calibration : 3-Mar-2026
Suggested calibration Due Date : 2-Mar-2027
Calibrated at : In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-03

Equipment Under Calibration

Physical Condition : Good

Name : Glass Thermometer
Make : GH Zeal
Model : --
S.No : --
ID No. : 22B

Range : -10 to 50 °C
Resolution : 0.5 °C
Accuracy : As Per Manual
Location : Microbiology Lab

Environmental Conditions

Temperature : 25.5 °C

Relative Humidity : 61.4 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	-10.0	-10.03	-10.02	-10.03	-10.03	0.03	--	0.40
2	0.0	-0.07	-0.08	-0.08	-0.08	0.08	--	0.36
3	10.0	9.89	9.88	9.89	9.89	0.11	--	0.36
4	30.0	29.80	29.79	29.79	29.79	0.21	--	0.36
5	50.0	49.60	49.62	49.62	49.61	0.39	--	0.36

** End of Certificate **

Checked By
(Harpriya. KH)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

