



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 03

ULR Number : CC2519 26 0 00002808F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,

A2 & A3 SIDCO Industrial Estate,

Andipalayam Post, Tiruchengode,

Namakkal - 637 214.

Date of Issue : 5-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0221

Receipt Date : 3-Mar-2026

Date of Calibration : 3-Mar-2026

Suggested calibration Due Date : 2-Mar-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-03

Equipment Under Calibration

Physical Condition : Good

Name : Glass Thermometer

Make : GH Zeal

Model : --

S.No : --

ID No. : 23

Range : -10 to 50 °C

Resolution : 0.5 °C

Accuracy : As Per Manual

Location : Microbiology Lab

Environmental Conditions

Temperature 25.5 °C

Relative Humidity 61.6 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	-10.0	-10.05	-10.05	-10.05	-10.05	0.05	--	0.40
2	0.0	-0.09	-0.09	-0.08	-0.09	0.09	--	0.36
3	10.0	9.90	9.91	9.90	9.90	0.10	--	0.36
4	30.0	29.82	29.83	29.82	29.82	0.18	--	0.36
5	50.0	49.71	49.70	49.71	49.71	0.29	--	0.36

** End of Certificate **



Checked By
(Signature)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

