



# EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



## CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 01

Date of Issue : 5-Mar-2026

ULR Number : CC2519 26 0 00002806F

Pages : 1 of 1

Customer Name and Address

SRF Number : 2026/0221

M/s. Christy Friedgram Industry Pvt Ltd,

Receipt Date : 3-Mar-2026

A2 & A3 SIDCO Industrial Estate,

Date of Calibration : 3-Mar-2026

Andipalayam Post, Tiruchengode,

Suggested calibration Due Date : 2-Mar-2027

Namakkal - 637 214.

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-03

### Equipment Under Calibration

Physical Condition : Good

Name : Glass Thermometer

Make : GH Zeal

Range : -50 to 50 °C

Model : --

Resolution : 1 °C

S.No : --

Accuracy : As Per Manual

ID No. : 25

Location : Microbiology Lab

### Environmental Conditions

Temperature 25.6 °C

Relative Humidity 61.6 %RH

Temperature Scale : ITS 90

### Standard Equipment used for Calibration

| S.No | Standard Instrument                    | Serial No             | Certificate number  | Validity   |
|------|----------------------------------------|-----------------------|---------------------|------------|
| 1    | RTD Sensor with 6 1/2 Digit Multimeter | 25000032 / TW00022681 | HT/CC/251213-20/001 | 12.12.2026 |

### Calibration Results

#### Parameter - Thermal/Temperature

| S.No | UUC Reading<br>°C | Standard Reading in °C |        |        | Standard Reading (Mean)<br>°C | Error<br>°C | Error Allowed<br>±°C | Uncertainty<br>±°C |
|------|-------------------|------------------------|--------|--------|-------------------------------|-------------|----------------------|--------------------|
|      |                   | 1                      | 2      | 3      |                               |             |                      |                    |
| 1    | -30               | -30.03                 | -30.03 | -30.03 | -30.03                        | 0.03        | --                   | 0.64               |
| 2    | 0                 | -0.09                  | -0.09  | -0.08  | -0.09                         | 0.09        | --                   | 0.62               |
| 3    | 10                | 9.81                   | 9.82   | 9.82   | 9.82                          | 0.18        | --                   | 0.62               |
| 4    | 30                | 29.76                  | 29.75  | 29.76  | 29.76                         | 0.24        | --                   | 0.62               |
| 5    | 50                | 49.51                  | 49.51  | 49.52  | 49.51                         | 0.49        | --                   | 0.62               |

\*\* End of Certificate \*\*



Remarks : See backside of this page

Authorized Signatory  
K. ABDUL VAJITH



ECT