



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 04
ULR Number : CC2519 26 0 00002809F
Customer Name and Address
M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 5-Mar-2026
Pages : 1 of 1
SRF Number : 2026/0221
Receipt Date : 3-Mar-2026
Date of Calibration : 3-Mar-2026
Suggested calibration Due Date : 2-Mar-2027
Calibrated at : In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-03

Equipment Under Calibration

Physical Condition : Good

Name : Glass Thermometer
Make : Vision
Model : --
S.No : --
ID No. : 28

Range : -10 to 110 °C
Resolution : 1 °C
Accuracy : As Per Manual
Location : Microbiology Lab

Environmental Conditions

Temperature : 25.3 °C

Relative Humidity : 61.5 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor with 6 1/2 Digit Multimeter	25000032 / TW00022681	HT/CC/251213-20/001	12.12.2026

Calibration Results

Parameter - Thermal/Temperature

S.No	UUC Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	-10	-10.03	-10.03	-10.03	-10.03	0.03	--	0.64
2	0	-0.08	-0.07	-0.07	-0.07	0.07	--	0.62
3	10	9.91	9.90	9.91	9.91	0.09	--	0.62
4	50	49.79	49.80	49.80	49.80	0.20	--	0.62
5	110	109.54	109.53	109.54	109.54	0.46	--	0.62

** End of Certificate **

Checked By
(Hampiya.KN)



Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

