



EURO CALIBRATION (OPC) PRIVATE LIMITED

No.387, IIInd Floor, TIC Road, Varadharajapuram, Ambattur, Chennai - 600 053.

Tamil Nadu. Tel : 044 - 4852 1288. Mobile : 9486195184, 9176032184.

E-mail : eurocalibration@gmail.com web : www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2026/0221- 15

ULR Number : CC2519 26 0 00002820F

Customer Name and Address

M/s. Christy Friedgram Industry Pvt Ltd,
A2 & A3 SIDCO Industrial Estate,
Andipalayam Post, Tiruchengode,
Namakkal - 637 214.

Date of Issue : 5-Mar-2026

Pages : 1 of 1

SRF Number : 2026/0221

Receipt Date : 3-Mar-2026

Date of Calibration : 3-Mar-2026

Suggested calibration Due Date : 2-Mar-2027

Calibrated at In House

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-12

Equipment Under Calibration

Physical Condition : Good

Name : Digital Thermo Hygrometer

Make : HTC

Model : HTC-1

S.No : --

ID No. : 31

Range : -10 to 50°C & 10 to 99% RH

Resolution : 0.1°C & 1% RH

Accuracy : $\pm 1^\circ\text{C}$ & 5 % RH

Location : Microbiology

Environmental Conditions

Temperature 25.4 °C

Relative Humidity 57.5 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Temperature & Humidity meter	T.009088	WOI/2025/2412/1	13.07.2026

Calibration Results

Parameter - Thermal / Specific Heat & Humidity

1. Temperature @ 50%RH

S.No	Standard Reading °C	UUC Reading in °C					UUC Reading (Mean) °C	Error °C	Error Allowed $\pm^\circ\text{C}$	Uncertainty $\pm^\circ\text{C}$
		1	2	3	4	5				
1	10.3	10.2	10.1	10.1	10.1	10.1	10.12	-0.18	1.00	0.75
2	27.7	27.4	27.3	27.4	27.3	27.3	27.34	-0.36	1.00	0.75
3	49.9	49.1	49.2	49.2	49.1	49.1	49.14	-0.76	1.00	0.75

2. Humidity @ 25 °C

S.No	Standard Reading % RH	UUC Reading in %RH					UUC Reading (Mean) % RH	Error % RH	Error Allowed $\pm\% \text{RH}$	Uncertainty $\pm\% \text{RH}$
		1	2	3	4	5				
1	35.6	35	35	35	35	35	35.00	-0.60	5.00	2.42
2	60.5	59	59	59	59	59	59.00	-1.50	5.00	2.42
3	93.5	90	90	90	90	90	90.00	-3.50	5.00	2.42

** End of Certificate **



Checked By
(Signature)

Remarks : See backside of this page

Authorized Signatory
K. ABDUL VAJITH

